

Certificate-Register-No.: Z - 08 - 362 - MP - E

Berlin Cert GmbH
Institute for Testing and Certification of Medical Devices
associated with the Technical University Berlin

certifies, that the following product of the company

Characteristics: Relative pressure relief compared to reference: 15%

The relative amplitude damping (shock absorbcency)
exceeds 50% (measured: $69,5 \pm 0,4\%$).

Characteristics: Relative pressure relief compared to reference: 15%

The relative amplitude damping (shock absorbcency)
exceeds 50% (measured: $69,5 \pm 0,4\%$).

The conducted tests are documented in the test protocol
P-08-362-MP-PA118-1-E and P-08-362-MP-PA118-2-E.

This certificate is not a confirmation, that the product complies with the German medical
device law or the directive 93/42/EEC

Issued on: 2008-09-12

